
Technical progress - detector performance, calibration, R&D efforts

Albrecht Karle March 2019


The IceCube Neutrino Observatory IceTop (surface array): 81 stations



IceCube: 86 strings 5160 optical sensors over 1 km³ volume 17 m vertical spacing 125 m horizontal spacing

Highly stable operation. Since 2016: livetime > 99.5%

DeepCore (low energy threshold)





South Pole 10m Telescope

IceCube Laboratory (ICL)

1 Aug




MAPO

TOS - Drilling site (79 & 80 in 10/11)

IceCube Enhanced Hot Water Drill (EHWD)


Photo: Ben Tibbets ~2009



to a depth between 1500 and 2500m.




DOM survival rate




DOMs display very high reliability. Only ~1 failure per year over the last 5 years.

Detector Uptime




IC86-2018 Cumulative IceCube Detector Time Usage



98.5% C 0.97% @ 0.32% E 0.22% C

PMT gain stability 2011 - 2016


No indication for any changes since 2016.




DOM gain appears stable!

(PMT gain of 1E7 is small.
Noise rates are small.
→ Very small integrated
current on anode.
→ No aging from that.

Time difference between



Understanding the ice



Measurement of South Pole ice transparency with the IceCube LED

calibration system,

Aartsen et al., (IceCube Coll.), NIMA55353 http://arxiv.org/abs/1301.5361


-500

20 -


40 -

Less scattering in direction of ice flow: \rightarrow up to ~10% /100m variation in amplitude


2. Azimuthal variation in of scattering



3. Ice layers are tilted – not planar










Systematic uncertainties: DOM and local ice



We plan to map the full surface sensitivity of every

Images taken with camera ("Swedish Camera") during refreeze process:




DOM and local ice


Hole ice visible on the right. Need to determine the effect for every single DOM.

Cable shadow

Cable diameter: 4.5cm




Azimuthal DOM response: Simulated effect on receiving DOM from flashers at close distance.




Built-in inclinometers vs DOM tilt fit


Indication of real tilt for 2 DOMs (out of 48)!

4 dozen DOMs have a built-in inclinometer, mounted on the mainboard, most of them have measured very small tilts, while 2 have tilts in excess of 20 degrees.





Example of DOM level calibration work: determined position of individual cables near DOM to few degree precision



azimuth to led7

40		
50		
60		
70		
80	المعادلات المحمد ا	

string

....

Types of events and interactions

Charged-current v_{μ} (data)

Up-going (throughgoing) track


Factor of ~2 energy resolution ~ 0.5° angular resolution


0.3° above 100 TeV

Isolated energy deposition (cascade) with no track

15% deposited energy resolution10-15° angular resolution (above 100 TeV)Working on improving that.


Early






Neutral-current / v_e

Charged-current v $_{\tau}$



(simulation)



"Double-bang"


(none observed yet: τ decay length is 50 m/ PeV)


> ID: above~ 100 TeV (two methods)

Bright DOMs

DOMs with Q_{bright} > 10*Q_{avg} are classified as "Bright"


PMT not necessarily saturated, but excluded because unmodeled systematic uncertainties start to dominate at high photon statistics







Local effects: DOM orientation and cable position


Without local effects








Observation of a 6 PeV neutrino of special interest



Typical visible energy is 93%


Work in progress



Event identified in a partially-contained PeV search (PEPE) Deposited energy: 5.9±0.18 PeV (stat only) 10¹⁸ ICRC 2017 arXiv:1710.01191

Energy resolution limited by systematic errors: Impact on science


More precise energy reconstruction \rightarrow



resolution on deposited energy

Statistical error: ~0.18 PeV Systematic error: ~0.7 PeV


reduces backgrounds for GR Reconstructed width of resonance (and thus background) increases by factor 4 due to sys. errors



Angular resolution of cascades: limited by systematics

Currently, reconstruction does not use the most nearby sensors in construction.

Bright signals don't help because of systematic errors. For simulated 1 PeV cascade \rightarrow Room to improve if local-ice is well-modeled




Identical ice model for simulation and reconstruction



Higher level performance parameters

Angular resolution for muon neutrinos

2019




Moon shadow

Cosmic rays absorbed by the moon result in a deficit of muons in IceCube


Moon, IC86-I 0.04 Relative Deficit З δ_{μ} - δ_{M} [deg] 0.02 -0.02 -0.04 0 -0.06 -0.08 -0.1 -2 -0.12 $(\alpha_{\mu} - \alpha_{M}) \cos(\delta_{\mu}) [deg]$ <u>–</u>3 -2 0

Detection of the Temporal Variation of the Sun's Cosmic Ray Shadow with the IceCube Detector ApJ, 872, (2019) 133

Absolute pointing verified to 0.1 ° Median resolution for low energy muons: 0.7°



Technical progress: TXS alert published 43 seconds after interaction.




Multimessenger astronomy in real time - flares Implementation of efficient realtime system online

Anna Franckowiak

- R&D related to M&O and continued optimization of IceCube:
 - Surface instrumentation
 - SpiceCore
- R&D geared towards the future: Upgrade and Gen2
 - Detector R&D, new optical modules




Snow depth of IceTop & effects on physics analysis



Snow accumulates on top of IceTop tanks at an average rate of 20 cm/year.

 >70% tanks are under 2 meters of snow or more. -----> Uncertainty affects a number of physics analyses



Science case for scintillator deployment

Enhance IceCube's neutrino measurements:

- Better understanding of atmospheric backgrounds from cosmic rays.
- Improved calibration of in-ice detectors.
- More efficient veto of cosmic ray backgrounds verification of crucial self veto method in energy range 10 to 100 TeV. The energy threshold at which the veto becomes efficient is estimated to be lower by a factor of two.


Cosmic Ray science

• More accurate measurements of the cosmic rays mass composition and energy spectrum above 1 PeV.


Other benefits: R&D for future detector upgrades

- A new, scalable precision timing and high-speed communications scheme for IceCube M&O and possible future projects.
- Efficient trenching procedures for instrumentation installation.
- Mechanical solutions to raise scintillator panels above the snow during the period of array deployment.

Scintillator deployment



Layout is optimized both for science and ease of deployment




Performance

- •Two stations (with different designs, each offering unique advantages) were installed in pole season 2018/19 and have been taking data since May 2018.
- •Excellent performance of both stations



0.00125 studo 0.00075 0.00050 0.00025 0.00000

Counts vs Charge - Chan 7




Radio component

3 radio antennas per scintillator station will help reconstructing gamma ray showers (which are particle-poor at surface) also from larger zenith angles (including the galactic center) and allow for:

- improving accuracy in the cosmic rays mass composition analysis
- discovery potential for PeV photons in a scenario, where the HESS source of the Galactic Center would be the source of the most energetic CR in the Milky Way
- testing hadronic interaction models

Science case backed up by simulation

Photo: Surface detector R&D January 2019





Air Cherenkov R&D

Measuring directly the CR showers electromagnetic component down to 20 TeV would:


- \rightarrow support calibration of the in-ice detector and IceTop
- \rightarrow Lower veto threshold for IceCube for the area covered.
- \rightarrow Improve mass composition analysis

Some challenges to be overcome (duty cycle, snow drift, electronics)

- 2 more IceAct telescopes at the South Pole are being installed in 2018/19 season
 - One will replace current telescope on ICL roof
 - One will be installed next to Scintillator Station, using connections at the scintillators Field Hub

Photo:

Surface detector R&D January 2019




Spice-Core: measuring ice in a borehole at the Pole

SpiceCore is a 4 inch bore hole left from an earlier ice-coring program: depth ~1700m distance: 2 km


ICNO:


Use this hole to deploy sensors and measure properties of the ice, such as: absorption in UV, radio propagation parameters



Spice-Core: measuring ice in a borehole at the Pole

Measurements yielded useful results and experience. Another measurement is planned for 2017/18





POCAM Integrating sphere

LED Pulser (250 nm, 285 nm,

3 mini-WOMs






7 strings in center of IceCube, densely instrumented

Science goals:

- v_{μ} disappearance
- v_{τ} appearance
- Precise calibration of IceCube optical properties and DOM response

A big step towards IceCube-Gen2

IceCube Upgrade (a step towards Gen2)



IceCube-Gen2 The next Generation IceCube: from discovery to astronomy

Multi-component observatory:

- IceCube-Gen2 High-Energy Array
- Surface air shower detector
- Sub-surface radio detector


Surface Area: ~6.5km² (0.9) Instrumented depth: 1.26 km (1.0)

Instrumented Volume: 8 km³

Order of magnitude increase of contained event rate at high energies.

> Artist conception Here: 120 strings at 300 m spacing






Sensor design R&D for improved performance

IceCube DOM





mDOM





33 cm


Directional information

36

- More sensitive area per module
- Directional information


30


- More sensitive area per module
- Smaller geometry



- more sensitive
 area per \$
 - Small diameter
- Lower noise rate
- Directional info.
- More area per module

FOM





Take away messages

- IceCube continues to evolve through improvements in understanding of ice, sensors and backgrounds that far exceed those anticipated in 2004.
- This knowledge results in improvements in performance, such as angular resolution that directly affects the results of multiple analyses.
- Systematic errors at all levels are increasingly important and vigorous efforts are underway to reduce them.
- Maintenance and R&D efforts such as surface instrumentation and measurements with in the SpiceCore will produce useful information.
- Detector R&D, sensor development, interface support is also happening to support the IceCube upgrade and maintain the ICNO facility as a support infrastructure for the future.

